

Product Recall.

Durex Real Feel Condoms 12 pack Durex Real Feel Condoms 18 pack

Date of recall: 31 July 2018

Reckitt Benckiser Group Plc has identified that a limited number of batches of the above condoms are not passing their stringent shelf-life durability tests as they fail the requirements for burst pressure towards the end of their shelf life.

Products recalled

Description	Batch code	Expiry date
Durex Real Feel 12 Pack	1000444370	Feb 2021
Durex Real Feel 18 Pack	1000430479	Jan 2021

Batch code information can be located on the bottom of the pack. On individual condoms, you will find the batch code on the back of the foil wrapping.

What you should do

Please do not use the affected batch codes of this product. Customers can also return the affected products to a Tesco store where a full refund will be given. No receipt is required. Use our store locator to [find your nearest Tesco](#).

Contact details

More information on returns can be found on the Durex website: www.durex.co.uk and www.durex.ie.

If you require further details, please contact Tesco Customer Services directly on [0800 505 555](tel:0800505555) (UK) and [1850 744 844](tel:1850744844) (ROI).

