

Product Recall.

Bausch + Lomb – selected products

Date of product recall: 10/06/21

We've been made aware that Bausch + Lomb are recalling 5 of their products, as the containers haven't undergone the correct sterilization process.



Product batches affected

All date codes of the products listed below are being recalled.

Products in the UK and ROI

Boston Advance Conditioner 120ml – product lot number: 53822876

Boston Advance Cleaner 30ml – product lot number: 53822398

Products in the UK only

Boston Simplus 120ml – product lot number: 72619858

ReNu Multiplus 240ml – product lot number: 53636941

Biotrue 300ml – product lot number: 66684419

What you should do

Visit www.blrecall.expertinquiry.com and do the following:

1. Insert your product lot number to check if your product is affected.
2. If your product is affected, please stop using it.
3. Follow the instructions to register your product.
4. Once done, please throw away the affected product.

Alternatively, if you can do so safely, you can return the product to a Tesco store where a full refund will be given. No receipt is required. Use our store locators to [find your nearest UK Tesco](#) or [nearest ROI Tesco](#).

If you're unable to come in-store, you can call us to arrange a refund.

Product Recall.

Bausch + Lomb – selected products

Date of product recall: 10/06/21

Contact details

We apologise for any inconvenience caused. If you would like further details, please contact Tesco Customer Services on [0800 505 555](tel:0800505555) (UK) and [1800 248 123](tel:1800248123) (ROI).

